



South Side

**WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION**

DATE: July 27, 2005 INITIALS: DEC OWNER: WSP
DISTRICT: 6 ADDRESS: I-90 West MP 80
LOCATION: Life Elem #53 COUNTY: Kittitas
MAKE: ISB TYPE: truck scale ACCURACY CLASS: TTL
MODEL: 651 ID. NO: 202173 CAPACITY: 200,000 lbs SECTIONS: 2
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: electronic PLATFORM: 12' X 20'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Platform</u>		<u>101</u>		<u>Platform</u>		<u>102</u>	
<u>Sect 1</u>	<u>20,000</u>	<u>19980</u>	<u>-20</u>	<u>Sect 1</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Center</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>	<u>Center</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Sect 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>	<u>Sect 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 22205-1 and 2

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronin

Date: 8/2/2005

State of Washington


County of Thurston

Signed and sworn to (or affirmed) before me on August 2, 2005 by

Michael J. Evans
(Signature of notary public)

Printed Name of Notary

My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS